

Certificate



No.: 968/K 204.05/23

Product tested	Seismic Monitoring System	Certificate holder	SYSCOM Instruments SA Rue de l'Industrie 21 1450 Sainte-Croix Switzerland
-----------------------	---------------------------	---------------------------	--

Type designation	NCC 2002, MR2002-SM24-K with sensor type MS2008+		
-------------------------	--	--	--

Codes and standards	IEC 62138:2018	IEC/IEEE 60780-323:2016	
----------------------------	----------------	-------------------------	--

Intended application	The devices comply with the requirements of the relevant standards (Category C acc. to IEC 62138). The assessment was executed in accordance with IEC 60780. Therefore the device can be used in applications up to Category C acc. to IEC 61226.		
-----------------------------	---	--	--

Specific requirements	The instructions of the associated Installation and Operating Manual shall be considered.		
------------------------------	---	--	--

Valid until 2028-06-28

The issue of this certificate is based upon an examination, whose results are documented in Report No. 968/K 204.05/23 dated 2023-06-27.

This certificate is valid only for products which are identical with the product tested.

TÜV Rheinland Industrie Service GmbH
Bereich Automation
Funktionale Sicherheit
Am Grauen Stein, 51105 Köln

Köln, 2023-06-29

Certification Body Safety & Security for Automation & Grid

H. Miedl
Dipl.-Inf. Horst Miedl